

Chart 5: C<sub>2</sub>H<sub>4</sub> (ethylene) analyzers

INTERFERING GAS														
C <sub>2</sub> H <sub>4</sub> (ethylene) ANALYZER	acetone	acetylene	Cl <sub>2</sub>	CO	CO <sub>2</sub>	ethyl alcohol	ethylene glycol	EtO	Freon	glutar- aldehyde	H <sub>2</sub>	isopropyl alcohol	N <sub>2</sub> O	NH <sub>3</sub>
	300	4	15 [N]	8	>10 <sup>4</sup>	150	0.3	0.2	>10 <sup>4</sup>	1	500	0.4	>10 <sup>5</sup>	200

[N] = Negative interference