

## Chart 6: H<sub>2</sub>O<sub>2</sub> (hydrogen peroxide) analyzers

INTERFERING GAS												
H <sub>2</sub> O <sub>2</sub> ANALYZER	H <sub>2</sub>	Cl <sub>2</sub>	CO	CO <sub>2</sub>	ethyl alcohol	EtO	H <sub>2</sub> S	SO <sub>2</sub>	glutar- aldehyde	isopropyl alcohol	N <sub>2</sub> O	NH <sub>3</sub>
	>8000	6 [N]	>1000	>10 <sup>4</sup>	>500	>500	0.6	2	>500	>500	>10 <sup>4</sup>	150