

JTPD.E114363 - GAS AND VAPOR DETECTION EQUIPMENT FOR USE IN HAZARDOUS LOCATIONS

Gas and Vapor Detection Equipment for Use in Hazardous Locations

See General Information for Gas and Vapor Detection Equipment for Use in Hazardous Locations

INTERSCAN CORP

E114363

SUITE 110

4590 ISH DR

SIMI VALLEY, CA 93063 USA

Portable gas analyzers, intrinsically safe when used with four SAFT part no. 750 SC size 1/2 C cell nickel cadmium batteries rated 1.2 v, 750 mAh each and two size C cell alkaline batteries rated 1/5 v each, "4000 Series", Model(s) 414*, 415*, 416*, 417*, 418*, 420*, 424*, 428*, 434*, 436*, 454*

* - Followed by a one digit number.

Last Updated on 2018-02-23

The appearance of a company's name or product in this database does not in itself assure that products so identified have been manufactured under UL's Follow-Up Service. Only those products bearing the UL Mark should be considered to be Certified and covered under UL's Follow-Up Service. Always look for the Mark on the product.

UL permits the reproduction of the material contained in the Online Certification Directory subject to the following conditions: 1. The Guide Information, Assemblies, Constructions, Designs, Systems, and/or Certifications (files) must be presented in their entirety and in a non-misleading manner, without any manipulation of the data (or drawings). 2. The statement "Reprinted from the Online Certifications Directory with permission from UL" must appear adjacent to the extracted material. In addition, the reprinted material must include a copyright notice in the following format: "© 2019 UL LLC"